



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Organization of:

FIT Pacific, Inc. Tsukuba Technical Center

3-20-1 Tokodai, Tsukuba, Ibaraki, Japan 300-2635

*and hereby declares that the Organization is accredited in accordance with
the recognized International Standard:*

ISO/IEC 17025:2017

Whereby, technical competence has been confirmed for the associated scope supplement, in the fields of:

Mechanical Testing
(As detailed in the supplement)

Accreditation claims for conformity assessment activities shall only be made from the addresses referenced within this certificate and shall apply solely to those activities identified in the related scope.

This Accreditation is granted subject to the Accreditation Body rules governing the Accreditation referred to above, and the Organization hereby commits to observing and complying with those rules in their entirety.

For PJLA:

Tracy Szerszen
President

Initial Accreditation Date:
May 28, 2011

Issue Date:
April 30, 2026

Expiration Date:
June 30, 2028

Accreditation No.:
69511

Certificate No.:
L26-422

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle.

The validity of this certificate should be confirmed through the PJLA website: www.pjlabs.com

Perry Johnson Laboratory Accreditation, Inc. (PJLA)
755 West Big Beaver Road, Suite 1325, Troy, Michigan 48084



Certificate of Accreditation: Supplement

FIT Pacific, Inc. Tsukuba Technical Center

3-20-1 Tokodai, Tsukuba, Ibaraki, Japan 300-2635

Contact Name: Nobuo Edakawa Phone: 029-848-0331

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Mechanical	Hybrid-III 95 th , 50 th , 5 th , 10YO, 6YO, 3YO, CRABI Head Drop Test	Head Drop Test Stand (TS-1)	IIII Dummy Calibration Procedure (TT-ID-TM10001) On basis of: -49 CFR Part 572 – Subpart E, N, O, P, R -SAE User's Manual the Hybrid III 95th Test Dummy Engineering Aid (EA-26) - SAE User's Manual H III 10YO Test Dummy Engineering Aid(EA-34)	Accelerometer Acceleration: 0 m/s ² to 2 942 m/s ² (0 g to 300 g) Unimodal Oscillation: 0 % to 17 % Temperature: 19 °C to 26 °C Humidity: 10 % to 70 %	F4	F
Mechanical	Side Impact Dummy ES-1, ES-2/2re, SID-II s (SBL-C&D) Head Drop Test	Head Drop Test Stand (TS-1)	Euro SID Calibration Procedure (TT-ID-TM10002), SID-II s Calibration Procedure (TT-ID-TM10003) On basis of: -49 CFR(Part 572) –Subpart U, V -ECE Addendum 94:Regulation 95 -User's Manual-ES-2 50 th Percentile Side Impact Crash Test Dummy -EuroNCAP_TRANS-WP29	Accelerometer Acceleration: 147 m/s ² -1 471 m/s ² (15 g – 150 g) Unimodal Oscillation: 0 % - 15 % Temperature: 20.6 °C - 22.2 °C Humidity: 10 % to 70 %	F4	F



Certificate of Accreditation: Supplement

FIT Pacific, Inc. Tsukuba Technical Center

3-20-1 Tokodai, Tsukuba, Ibaraki, Japan 300-2635

Contact Name: Nobuo Edakawa Phone: 029-848-0331

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Mechanical	Hybrid-III 95 th , 50 th , 5 th , 10YO, 6YO, 3YO, CRABI Neck Extension Test	Neck Pendulum Test Stand (TS-2)	IIII Dummy Calibration Procedure (TT-ID-TM10001) On basis of: -49 CFR(Part 572) – Subpart E, N, O, P, R -SAE User's Manual IIII 95th Test Dummy Engineering Aid(EA-26) -SAE User's Manual IIII 10YO Test Dummy Engineering Aid(EA-34)	Phototube, Accelerometer, Angle Potentiometer, Load Cell Velocity: 1.20 m/s - 7.13 m/s Acceleration: 0 m/s ² - 284.4 m/s ² (0.0 g - 29.0 g) Time: 34 ms - 128 ms Angle: 64 ° - 92 ° Moment: 0 Nm - 130 Nm Temperature: 20.6 °C - 22.2 °C Humidity: 10 % - 70 %	F4	F



Certificate of Accreditation: Supplement

FIT Pacific, Inc. Tsukuba Technical Center

3-20-1 Tokodai, Tsukuba, Ibaraki, Japan 300-2635

Contact Name: Nobuo Edakawa Phone: 029-848-0331

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Mechanical	Hybrid-III 95th, 50th, 5th, 10YO, 6YO, 3YO, CRABI Neck Flexion Test	Neck Pendulum Test Stand (TS-2)	IIII Dummy Calibration Procedure (TT-ID-TM10001) On basis of: -49 CFR(Part 572) – Subpart E, N, O, P, R -SAE User's Manual IIII 95th Test Dummy Engineering Aid(EA-26) -SAE User's Manual IIII 10YO Test Dummy Engineering Aid(EA-34)	Phototube, Accelerometer, Angle Potentiometer, Load Cell Velocity: 1.20 m/s - 7.13 m/s Acceleration: 0 m/s ² - 284.4 m/s ² (0.0 g - 29.0 g) Time: 34 ms - 128 ms Angle: 64 ° - 92 ° Moment: 0 Nm - 130 Nm Temperature: 20.6 °C - 22.2 °C Humidity: 10 % - 70 %	F4	F



Certificate of Accreditation: Supplement

FIT Pacific, Inc. Tsukuba Technical Center

3-20-1 Tokodai, Tsukuba, Ibaraki, Japan 300-2635

Contact Name: Nobuo Edakawa Phone: 029-848-0331

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Mechanical	Side Impact Dummy ES-1, ES-2/2re Lumbar Flexion Test	Neck Pendulum Test Stand (TS-2)	Euro SID Calibration Procedure (TT-ID-TM10002) On basis of: -49 CFR(Part 572) – Subpart U -ECE Addendum 94: Regulation 95 -User's Manual – ES-2 50 th Percentile Side Impact Crash Test Dummy -Euro NCAP_TRANS-WP29	Phototube, Accelerometer, Angle Potentiometer, Load Cell Velocity: 5.95 m/s to 6.15 m/s Acceleration: 0 m/s ² to 334.4 m/s ² (0 g to 34.1 g) Deceleration: 0 m/s to 6.5 m/s Time: 0 ms to 31.8 ms Angle: 0° to 55° Moment: 37 Nm to 57 Nm Temperature: 20.6 °C to 22.2 °C Humidity: 10 % to 70 %	F4	F



Certificate of Accreditation: Supplement

FIT Pacific, Inc. Tsukuba Technical Center

3-20-1 Tokodai, Tsukuba, Ibaraki, Japan 300-2635

Contact Name: Nobuo Edakawa Phone: 029-848-0331

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Mechanical	Side Impact Dummy ES-1, ES-2/2re, SID-II s (SBL-C&D) Neck Flexion Test	Neck Pendulum Test Stand (TS-2)	Euro SID Calibration Procedure (TT-ID-TM10002), SID-II s Calibration Procedure (TT-ID-TM10003) On basis of: -49 CFR Part 572 – Subpart U, V -ECE Addendum 94: Regulation 95 -User's Manual – ES-2 50 th Percentile Side Impact Crash Test Dummy -Euro NCAP_TRANS-WP29	Phototube, Accelerometer, Angle Potentiometer, Load Cell Velocity: 3.3 m/s to 5.63 m/s Acceleration: 0 m/s ² to 362.8 m/s ² (0 g to 37 g) Deceleration: 0 m/s to 6.4 m/s Time: 0 ms to 60 ms Angle: 46° to 82° Moment: 36 Nm to 44 Nm Temperature: 20.6 °C to 22.2 °C Humidity: 10 % to 70 %	F4	F



Certificate of Accreditation: Supplement

FIT Pacific, Inc. Tsukuba Technical Center

3-20-1 Tokodai, Tsukuba, Ibaraki, Japan 300-2635

Contact Name: Nobuo Edakawa Phone: 029-848-0331

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Mechanical	Hybrid-III 95 th , 50 th , 5 th , 10YO, 6YO Knee Impact Test	Knee Impact Test Stand (TS-3)	HIII Dummy Calibration Procedure (TT-ID-TM10001) On basis of: -49 CFR(Part 572) — Subpart E, N, O -SAE User's Manual H III 95 th Test Dummy Engineering Aid(EA-26) -SAE User's Manual H III 10YO Test Dummy Engineering Aid(EA-34)	Phototube, Accelerometer, Velocity: 2.07 m/s to 2.13 m/s Force: 2.0 N to 6.0 N Temperature: 19 °C to 26 °C Humidity: 10 % to 70 %	F4	F
Mechanical	Hybrid-III 95 th , 50 th , 5 th Knee Slider Test	Knee Impact Test Stand (TS-3)	HIII Dummy Calibration Procedure (TT-ID-TM10001) On basis of: -SAE User's Manual HIII 50 th Test Dummy Engineering Aid(EA-23) -SAE User's Manual HIII 5 th Test Dummy Engineering Aid(EA-25) -SAE User's Manual HIII 95 th Test Dummy Engineering Aid (EA-26) -J2876 HIII 50th Low Speed Ball Knee Slider Test	Phototube, Load Cell, Linear Potentiometer Velocity: 2.70 m/s to 2.80 m/s Force: 1.26 kN to 3.10 kN Displacement: 10 mm to 18.3 mm Temperature: 19 °C to 26 °C Humidity: 10 % to 70 %	F4	F



Certificate of Accreditation: Supplement

FIT Pacific, Inc. Tsukuba Technical Center

3-20-1 Tokodai, Tsukuba, Ibaraki, Japan 300-2635

Contact Name: Nobuo Edakawa Phone: 029-848-0331

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Mechanical	Hybrid-III 95 th , 50 th , 5 th , 10YO, 6YO, 3YO, CRABI Thorax Impact Test	Thorax Impact Test Stand (TS-4)	HIII Dummy Calibration Procedure (TT-ID-TM10001) On basis of: -49 CFR(Part 572) — Subpart E, N, O, P, R -SAE User's Manual H III 95th Test Dummy Engineering Aid(EA-26) -SAE User's Manual H III 10YO Test Dummy Engineering Aid(EA-34) -J2878 HIII 5th Low Speed Thorax Impact Test	Phototube, Accelerometer, Linear Potentiometer Velocity: 4.90 m/s to 6.83 m/s Force: 680 N to 5 900 N Displacement: 12.5 mm to 76.0 mm Hysteresis: 65 % to 85 % Temperature: 20.6 °C to 22.2 °C Humidity: 10 % to 70 %	F4	F
Mechanical	Side Impact Dummy ES-1, ES-2/2re, SID-II s (SBL-C&D) Shoulder Impact Test	Thorax Impact Test Stand (TS-4)	Euro SID Calibration Procedure (TT-ID-TM10002) SID-II s Calibration Procedure (TT-ID-TM10003) On basis of: -49 CFR(Part 572)— Subpart U, V -ECE Addendum 94: Regulation 95 -User's Manual – ES-2 50th Percentile Side Impact Crash Test Dummy -Euro NCAP_TRANS-WP29	Phototube, Accelerometer, Linear Potentiometer Velocity: 4.20 - 4.59 m/s Acceleration: 73.5 - 186.3 m/s ² (7.5 -19 g) Force: 1.6 - 2.4 kN Displacement: 27- 39 mm Temperature: 20.6 - 22.2 °C Humidity: 10 - 70 %	F4	F



Certificate of Accreditation: Supplement

FIT Pacific, Inc. Tsukuba Technical Center

3-20-1 Tokodai, Tsukuba, Ibaraki, Japan 300-2635

Contact Name: Nobuo Edakawa Phone: 029-848-0331

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Mechanical	Side Impact Dummy ES-1, ES-2/2re, SID-II s (SBL-C&D) Thorax Impact Test (With Arm and Without Arm)	Thorax Impact Test Stand (TS-4)	Euro SID Calibration Procedure (TT-ID-TM10002), SID-IIs Calibration Procedure (TT-ID-TM10003) On basis of: -49 CFR(Part 572)— Subpart U, V -ECE Addendum 94: Regulation 95 -User's Manual – ES-2 50 th Percentile Side Impact Crash Test Dummy -Euro NCAP_TRANS-WP29	Phototube, Accelerometer, Linear Potentiometer Velocity: 4.20 m/s - 6.84 m/s Acceleration: 68 m/s ² - 470 m/s ² (7 g - 48 g) Force: 1.8 kN - 6.2 kN Displacement: 23 mm - 51 mm Temperature: 20.6 °C - 22.2 °C Humidity: 10 % - 70 %	F4	F



Certificate of Accreditation: Supplement

FIT Pacific, Inc. Tsukuba Technical Center

3-20-1 Tokodai, Tsukuba, Ibaraki, Japan 300-2635

Contact Name: Nobuo Edakawa Phone: 029-848-0331

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Mechanical	Side Impact Dummy ES-1, ES-2/2re, SID-II s (SBL-C&D) Abdominal Impact Test	Thorax Impact Test Stand (TS-4)	Euro SID Calibration Procedure (TT-ID-TM10002), SID-IIs Calibration Procedure (TT-ID-TM10003) On basis of: -49 CFR(Part 572)— Subpart U, V -ECE Addendum 94: Regulation 95 -User's Manual – ES-2 50 th Percentile Side Impact Crash Test Dummy -Euro NCAP_TRANS-WP29	Phototube, Accelerometer, Load Cell, Linear Potentiometer Velocity: 3.9 m/s to 6.4 m/s Acceleration: 49 m/s ² to 157 m/s ² (5 g to 16 g) Force: 1.3 kN to 9.5 kN Time: 8.5 ms to 13 ms Displacement: 33 mm to 56 mm Temperature: 20.6 °C to 22.2 °C Humidity: 10 % to 70 %	F4	F
Mechanical	Side Impact Dummy ES-1, ES-2/2re, SID-II s (SBL-C&D) Pelvis Impact Test	Thorax Impact Test Stand (TS-4)	Euro SID Calibration Procedure (TT-ID-TM10002), SID-IIs Calibration Procedure (TT-ID-TM10003) On basis of: -49 CFR(Part 572)— Subpart U, V -ECE Addendum 94: Regulation 95 -User's Manual – ES-2 50 th Percentile Side Impact Crash Test Dummy -Euro NCAP_TRANS-WP29	Phototube, Accelerometer, Load Cell Velocity: 4.2 m/s to 6.84 m/s Acceleration: 333 m/s ² to 588 m/s ² (34 g to 60 g) Force: 1.04 kN to 7.2 kN Time: 9.9 ms to 15.9 ms Temperature: 20.6 °C to 22.2 °C Humidity: 10 % to 70 %	F4	F



Certificate of Accreditation: Supplement

FIT Pacific, Inc. Tsukuba Technical Center

3-20-1 Tokodai, Tsukuba, Ibaraki, Japan 300-2635

Contact Name: Nobuo Edakawa Phone: 029-848-0331

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Mechanical	Side Impact Dummy SID-II s (SBL-D) Iliac Wing Impact Test	Thorax Impact Test Stand (TS-4)	SID-IIs Calibration Procedure (TT-ID-TM10003) On basis of: -49 CFR(Part 572) – Subpart V	Phototube, Accelerometer, Load Cell, Velocity: 4.2 m/s to 4.4 m/s Acceleration: 274 m/s ² to 441 m/s ² (28 g to 45 g) Force: 4.1 kN to 5.1 kN Temperature: 20.6 °C to 22.2 °C Humidity: 10 % to 70 %	F4	F
Mechanical	BioRID-II Mini Sled Test	Thorax Impact Test Stand (TS-4)	Bio RID-II Dummy Calibration Procedure (TT-ID-TM10004) On basis of: Denton BioRID-II User's Manual	Phototube, Accelerometer, Angle Potentiometer Velocity: 4.7 m/s to 4.8 m/s Acceleration: 117.6 m/s ² to 1 755.4 m/s ² (12.0 g to 179.0 g) Force: 4.70 kN to 5.20 kN Time: 4.0 ms to 24.0 ms Temperature: 19.0 °C to 25.0 °C Humidity: 10 % to 70 %	F4	F



Certificate of Accreditation: Supplement

FIT Pacific, Inc. Tsukuba Technical Center

3-20-1 Tokodai, Tsukuba, Ibaraki, Japan 300-2635

Contact Name: Nobuo Edakawa Phone: 029-848-0331

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Mechanical	Hybrid-III 95 th , 5 th , 10YO, 6YO, 3YO Torso Flexion Test	Torso Flexion Test Stand (TS-5)	HIII Dummy Calibration Procedure (TT-ID-TM10001) On basis of: -49 CFR(Part 572) — Subpart N, O, P -SAE User's Manual H III 95 th Test Dummy Engineering Aid (EA-26) -SAE User's Manual H III 10YO Test Dummy Engineering Aid(EA-34)	Load Cell, Angle Potentiometer Angular Velocity: 0.5 °/s to 1.5 °/s Force: 130 N to 550 N Angle: 0.0° to 45.5° Temperature: 20.6 °C to 22.2 °C Humidity: 10 % to 70 %	F4	F
Mechanical	Hybrid-III 50 th , 5 th Hip Range Motion Test	Hip Range Motion Test Stand (TS-6)	HIII Dummy Calibration Procedure (TT-ID-TM10001) On basis of: -49 CFR(Part 572) — Subpart E -SAE User's Manual H III 5 th Test Dummy Engineering Aid (EA-25)	Load Cell, Angle Potentiometer Angle: 0° to 50° Moment: 0 Nm to 203 Nm Angular Velocity: 5 °/s to 10 °/s Temperature: 19 °C to 26 °C Humidity: 10 % to 70 %	F4	F



Certificate of Accreditation: Supplement

FIT Pacific, Inc. Tsukuba Technical Center

3-20-1 Tokodai, Tsukuba, Ibaraki, Japan 300-2635

Contact Name: Nobuo Edakawa Phone: 029-848-0331

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Mechanical	Side Impact Dummy ES-1, ES-2/2re Damper Module Test	Drop Tower Test Stand (TS-7)	Euro SID Calibration Procedure (TT-ID-TM10002) On basis of: -49 CFR(Part 572)—Subpart U -ECE Addendum 94: Regulation 95 -User's Manual – ES-2 50 th Percentile Side Impact Crash Test Dummy -Euro NCAP_TRANS-WP29	Phototube, Accelerometer, Linear Potentiometer Velocity: 0.98 m/s to 4.10 m/s Displacement: 10.0 mm to 51.0 mm Temperature: 20.6 °C to 22.2 °C Humidity: 10 % to 70 %	F4	F
Mechanical	Hybrid-III 50 th – Eurofoot Lower Foot Impact Test	Foot Impact Test Stand (TS-8)	HIID Dummy Calibration Procedure (TT-ID-TM10001) On basis of: ECE Addendum 93: Regulation 94	Phototube, Accelerometer, Velocity: 4.3 m/s to 4.5 m/s Acceleration: 2 402 m/s ² to 3 383 m/s ² (245 g to 345 g) Temperature: 19 °C to 25 °C Humidity: 10 % to 70 %	F4	F



Certificate of Accreditation: Supplement

FIT Pacific, Inc. Tsukuba Technical Center

3-20-1 Tokodai, Tsukuba, Ibaraki, Japan 300-2635

Contact Name: Nobuo Edakawa Phone: 029-848-0331

Accreditation is granted to the facility to perform the following conformity assessment activities:

FIELD OF TEST	ITEMS, MATERIALS, OR PRODUCTS TESTED	COMPONENT, CHARACTERISTIC, PARAMETER TESTED	SPECIFICATION OR STANDARD METHOD	TECHNOLOGY OR TECHNIQUE USED	FLEX CODE	LOCATION OF ACTIVITY
Mechanical	Hybrid-III 50 th – Eurofoot Upper Foot Impact Test	Foot Impact Test Stand (TS-8)	HIII Dummy Calibration Procedure (TT-ID-TM10001) On basis of: ECE Addendum 93: Regulation 94	Phototube, Accelerometer, Load Cell Velocity: 6.6 m/s to 6.8 m/s Moment: 95 Nm to 145 Nm Temperature: 19 °C to 25 °C Humidity: 10 % to 70 %	F4	F
Mechanical	Hybrid-III 50 th – Eurofoot Lower Foot Impact with Shoe Test	Foot Impact Test Stand (TS-8)	HIII Dummy Calibration Procedure (TT-ID-TM10001) On basis of: ECE Addendum 93: Regulation 94	Phototube, Accelerometer, Load Cell Velocity: 6.6 m/s to 6.8 m/s Force: 2.8 N to 3.8 N Temperature: 19 °C to 25 °C Humidity: 10 % to 70 %	F4	F

1. Location of activity: Location Code – Location
F - Conformity assessment activity is performed at the CABs fixed facility
2. Flex Code:
F0- Fixed scope item. No deviations allowed to the line item as identified, except for updating to the most recent version of an accredited standard method after verification.
F1- Laboratory has the capability to test a new item, material, matrix, or product similar in composition to item, material, matrix, or product identified on the scope
F2- Laboratory has the capability to introduce the newest revision of an accredited authoritative standard method (with no modifications) identified on the scope
F3- Laboratory has the capability to introduce a parameter/component/analyte to an accredited test method identified on the scope
F4- Laboratory has the capability to introduce a new revision of an accredited non-standard method using the same technology or technique identified on the scope
F5- Laboratory has the capability to introduce a validated method that is equivalent to an accredited method (using same technology or technique) identified on the scope