



PERRY JOHNSON LABORATORY ACCREDITATION, INC.

Certificate of Accreditation

Perry Johnson Laboratory Accreditation, Inc. has assessed the Laboratory of:

FIT Pacific, Inc. Tsukuba Technical Center
3-20-1 Tokodai, Tsukuba, Ibaraki, Japan 300-2635

(Hereinafter called the Organization) and hereby declares that Organization is accredited in accordance with the recognized International Standard:

ISO/IEC 17025:2017

This accreditation demonstrates technical competence for a defined scope and the operation of a laboratory quality management system (as outlined by the joint ISO-ILAC-IAF Communiqué dated April 2017):

Dimensional Calibration, Mass, Force, and Weighing Devices Calibration, Acoustic Calibration, Electrical Calibration
(As detailed in the supplement)

Accreditation claims for such testing and/or calibration services shall only be made from addresses referenced within this certificate. This Accreditation is granted subject to the system rules governing the Accreditation referred to above, and the Organization hereby covenants with the Accreditation body's duty to observe and comply with the said rules.

For PJLA:

Tracy Szerszen
President

Initial Accreditation Date:

May 28, 2011

Issue Date:

May 24, 2024

Expiration Date:

June 30, 2026

Accreditation No.:

69511

Certificate No.:

L24-392

Perry Johnson Laboratory
Accreditation, Inc. (PJLA)
755 W. Big Beaver Rd., Suite 1325
Troy, Michigan 48084

The validity of this certificate is maintained through ongoing assessments based on a continuous accreditation cycle. The validity of this certificate should be confirmed through the PJLA website: www.pjlab.com



Certificate of Accreditation: Supplement

FIT Pacific, Inc. Tsukuba Technical Center
 3-20-1 Tokodai, Tsukuba, Ibaraki, Japan 300-2635
 Contact Name: Nobuo Edakawa Phone: 029-848-0331

Accreditation is granted to the facility to perform the following calibrations:

Dimensional

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Linear Potentiometer ^F Calibration Fixed Points Fixture (Model 5245)	0 mm 140 mm	$(0.49 + 3.5 \times 10^{-3}L)$ mm $(3.5 \mu\text{m}/\text{mm} + 0.49 \text{ mm})$	SAE 2517 Chest/Linear Pot. Calibration Procedure (TT-ID-TM10005) Linear Pot. Cal. Fixture Calibration Procedure (TT-ID-TM10007)
Rotary Potentiometer ^F Calibration through 150° Rotational Displacement Fixed Points Fixture (Model 6790)	0° 150°	$(0.52 + 3.5 \times 10^{-3}D)$ ° $0.0035 \text{ }^\circ/\text{ }^\circ + 0.52^\circ$	User's manual of Denton ATD, Inc. Rotary Pot. Calibration Procedure(TT-ID-TM10006) Rotary Pot. Cal. Fixture Calibration Procedure (TT-ID-TM10008)

Mass, Force, and Weighing Devices

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Load Cell ^F Axial Force Moment:	25 kN maximum	$(0.78 \times 10^{-3}F + 21)$ N	SAE J211-1 -Load Cell Calibration Procedure:NT-ID-TM001 -Output measurement for weight(mV/V)
Arm Length 50.8 mm		$(1.64 \times 10^{-3}F + 41)$ N	
Arm Length 38.1 mm		$(2.06 \times 10^{-3}F + 51)$ N	
Arm Length 31.75 mm		$(2.44 \times 10^{-3}F + 61)$ N	
Arm Length 25.4 mm		$(2.99 \times 10^{-3}F + 75)$ N	
Load Cell ^F Axial Force Moment:	10 kN maximum	$(0.78 \times 10^{-3}F + 8)$ N	SAE J211-1 -Load Cell Calibration Procedure: NT-ID-TM001 -Output measurement unit for weight: mV/V
Arm Length 50.8 mm		$(1.64 \times 10^{-3}F + 16)$ N	
Arm Length 38.1 mm		$(2.06 \times 10^{-3}F + 21)$ N	
Arm Length 31.75 mm		$(2.44 \times 10^{-3}F + 24)$ N	
Arm Length 25.4 mm		$(2.99 \times 10^{-3}F + 30)$ N	
			Calibration equipment: 1110AO-25KN
			Calibration equipment: SSM-AF-10KN



Certificate of Accreditation: Supplement

FIT Pacific, Inc. Tsukuba Technical Center
 3-20-1 Tokodai, Tsukuba, Ibaraki, Japan 300-2635
 Contact Name: Nobuo Edakawa Phone: 029-848-0331

Accreditation is granted to the facility to perform the following calibrations:

Acoustic

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED	
Acceleration sensor ^F	Voltage sensitivity: mV/(m/s ²) Acceleration meter output ≥ 0.01 mV		‘Calibration procedures for acceleration meter and angular rate sensor’ (NT-IA-TM001) On basis of: ISO 16063-21 (vibration type) Calibration device: CS18MF(SPEKTRA) PRBox (only voltage)	
	5 Hz to 100 Hz	1.3 %		
	100 Hz	1.3 %		
	100 Hz to 2 kHz	1.7 %		
	2 kHz to 5 kHz	1.7 %		
	Charge sensitivity: pC/(m/s ²) Acceleration meter output ≥ 1 pC			Calibration equipment: 301A10(PCB)
	5 Hz to 100 Hz	1.3 %		
	100 Hz	1.3 %		
	100 Hz to 2 kHz	1.7 %		
	Voltage sensitivity: mV/(m/s ²) Acceleration meter output ≥ 0.01 mV			‘Calibration procedures for acceleration meter and angular rate sensor’ (NT-IA-TM001) On basis of: ISO 16063-21 (vibration type) Calibration device: CS18MF(SPEKTRA) PRBox (only voltage)
	10 Hz to 100 Hz	1.3 %		
	100 Hz	1.3 %		
	100 Hz to 2 kHz	1.7 %		
	2 kHz to 5 kHz	1.7 %		
	5 kHz to 10 kHz	3.1 %		
	Charge sensitivity: pC/(m/s ²) Acceleration meter output ≥ 1 pC		Calibration equipment: 8305(B&K)	
	10 Hz to 100 Hz	1.3 %		
	100 Hz	1.3 %		
	100 Hz to 2 kHz	1.7 %		
	2 kHz to 5 kHz	1.7 %		
5 kHz to 10 kHz	3.1 %			



Certificate of Accreditation: Supplement

FIT Pacific, Inc. Tsukuba Technical Center
 3-20-1 Tokodai, Tsukuba, Ibaraki, Japan 300-2635
 Contact Name: Nobuo Edakawa Phone: 029-848-0331

Accreditation is granted to the facility to perform the following calibrations:

Acoustic

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (±)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Acceleration sensor ^F	Voltage sensitivity: mV/(m/s ²) Acceleration meter output ≥ 0.01 mV		‘Calibration procedures for acceleration meter and angular rate sensor’ (NT-IA-TM001) On basis of: ISO 16063-21 (vibration type) Calibration device: CS18HF(SPEKTRA) PRBox (only voltage) Calibration equipment: M353B17(PCB)
	5 Hz to 10 Hz	1.3 %	
	10 Hz to 20 Hz	1.1 %	
	20 Hz to 1 kHz	1.0 %	
	1 kHz to 5 kHz	1.1 %	
	5 kHz to 10 kHz	1.7 %	
	10 kHz to 15 kHz	2.2 %	
	15 kHz to 20 kHz	3.1 %	
	Charge sensitivity: pC/(m/s ²) Acceleration meter output ≥ 1 pC		
	5 Hz to 10 Hz	1.3 %	
	10 Hz to 20 Hz	1.1 %	
	20 Hz to 1 kHz	1.0 %	
	1 kHz to 5 kHz	1.1 %	
	5 kHz to 10 kHz	1.7 %	
	10 kHz to 15 kHz	2.2 %	
15 kHz to 20 kHz	3.1 %		
Angular rate sensor ^F 25 deg (p-p)	Charge sensitivity: mV/(deg/s) Angular rate sensor output ≥ 0.01 mV		‘Calibration procedures for acceleration meter and angular rate sensor’ (NT-IA-TM001) On basis of: ISO 16063-21 (vibration type) Calibration device: CS18HF(SPEKTRA) PRBox (only voltage) Calibration equipment: DTS ARS-PRO-8K
	1 Hz to 200 Hz	1.6 %	



Certificate of Accreditation: Supplement

FIT Pacific, Inc. Tsukuba Technical Center
3-20-1 Tokodai, Tsukuba, Ibaraki, Japan 300-2635
Contact Name: Nobuo Edakawa Phone: 029-848-0331

Accreditation is granted to the facility to perform the following calibrations:

Electrical

MEASURED INSTRUMENT, QUANTITY OR GAUGE	RANGE OR NOMINAL DEVICE SIZE AS APPROPRIATE	CALIBRATION AND MEASUREMENT CAPABILITY EXPRESSED AS AN UNCERTAINTY (\pm)	CALIBRATION EQUIPMENT AND REFERENCE STANDARDS USED
Acceleration sensor ^F	Voltage sensitivity: mV/(m/s ²) Acceleration meter output \geq 0.01 mV		‘Calibration procedures for acceleration meter and angular rate sensor’ (NT-IA-TM001) On basis of: ISO 16063-22 (Shock Type) Calibration device: CS18LS(SPEKTRA) PRBox (only voltage) Calibration equipment: 2270(ENDEVCO)
	98 m/s ² to 1960 m/s ²	0.8 %	
	Charge sensitivity: pC/(m/s ²) Acceleration meter output \geq 1 pC		
	98 m/s ² to 1960 m/s ²	0.8 %	

1. The CMC (Calibration and Measurement Capability) stated for calibrations included on this scope of accreditation represent the smallest measurement uncertainties attainable by the laboratory when performing a more or less routine calibration of a nearly ideal device under nearly ideal conditions. It is expressed at a confidence level of 95 % using a coverage factor *k* (usually equal to 2). The actual measurement uncertainty associated with a specific calibration performed by the laboratory will typically be larger than the CMC for the same calibration since capability and performance of the device being calibrated and the conditions related to the calibration may reasonably be expected to deviate from ideal to some degree.
2. The laboratories range of calibration capability for all disciplines for which they are accredited is the interval from the smallest calibrated standard to the largest calibrated standard used in performing the calibration. The low end of this range must be an attainable value for which the laboratory has or has access to the standard referenced. Verification of an indicated value of zero in the absence of a standard is common practice in the procedure for many calibrations but by its definition it does not constitute calibration of zero capacity.
3. The presence of a superscript F means that the laboratory performs calibration of the indicated parameter at its fixed location. Example: Outside Micrometer ^F would mean that the laboratory performs this calibration at its fixed location.